

## MicroAmp™ plastic consumables compatibility chart for Applied Biosystems™ endpoint PCR systems and genetic analyzers

Product	Cat. No.	3 x 32-well	96-well 0.2 mL		96-well 0.1 mL	384-well		Genetic analyzers		
		ProFlex™	ProFlex, SimpliAmp™, VeritiPro™, Veriti™, MiniAmp™ Plus, MiniAmp	2720	9700	Veriti	ProFlex, Veriti, VeritiPro™	9700	310	3130, 3130xl, 3500, 3500xl, 3730, 3730xl
<b>96-well 0.2 mL reaction plates</b>										
Optical 96-Well Plate	N8010560, 4316813		•	•	•					•
Optical 96-Well Plate with Barcode	4306737, 4326659		•	•	•					•
96-Well Plate with Barcode & Optical Caps	403012		•	•	•					
Optical 96-Well Plate with Barcode & Optical Adhesive Films	4314320		•	•	•					
EnduraPlate Optical 96-Well Clear Plate	A36924		•	•	•					•
EnduraPlate Optical 96-Well Clear Plate with Barcode*	4483354, 4483352		•	•	•					•
TriFlex 3 x 32-Well Reaction Plate	A32810, A32811	•	•	•	•					
<b>96-well 0.1 mL reaction plates</b>										
Fast Optical 96-Well Plate, 0.1 mL	4346907				•				•	•
Fast Optical 96-Well Plate with Barcode, 0.1 mL	4346906, 4366932				•				•	•
EnduraPlate Optical 96-Well Fast Plate	A36930				•				•	•
EnduraPlate Optical 96-Well Fast Clear Plate with Barcode*	4483485, 4483494				•				•	•
<b>384-well reaction plates</b>										
Optical 384-Well Plate	4343370					•	•			•
Optical 384-Well Plate with Barcode	4309849, 4326270, 4343814					•	•			•
EnduraPlate Optical 384-Well Plate	A36931					•	•			•
EnduraPlate Optical 384-Well Clear Plate with Barcode*	4483285, 4483273					•	•			•
<b>Strip tubes and caps</b>										
Fast 8-Tube Strip, 0.1 mL	4358293				•					
Optical 8-Tube Strip with Attached Optical Caps, 0.2 mL	A30588	•	•	•	•					
8-Tube Strip with Attached Domed Caps, 0.2 mL	A30589	•	•	•	•					
8-Tube Strip, 0.2 mL*	N8010580	•	•	•	•				•	
Optical 8-Tube Strip, 0.2 mL	4316567	•	•	•	•					
8-Cap Strip*	N8010535, N8011535	•	•	•	•	•				
Optical 8-Cap Strip	4323032	•	•	•	•	•				
12-Cap Strip*	N8010534, N8011534	•	•	•	•	•				
<b>Single tubes</b>										
Fast Reaction Tube with Cap, 0.1 mL	4358297, 4358293				•					
Reaction Tube with Cap, 0.2 mL*	N8010540, N8010612, N8011540	•	•	•	•					
Reaction Tube without Cap, 0.2 mL*	N8010533, N8011533	•	•	•	•					
Optical Tube without Cap, 0.2 mL	N8010933	•	•	•	•					
<b>Seals and covers</b>										
Clear Adhesive Film	4306311		•	•	•	•	•			
Optical Adhesive Film	4360954, 4311971		•	•	•	•	•			
96-Well Full Plate Cover	N8010550			•	•					
32-Well Clear Adhesive Film	A32812	•	•	•	•					
<b>Accessories</b>										
Splash-Free 96-Well Base	4312063		•	•	•	•				
96-Well Support Base	4379590		•	•	•	•				•
96-Well Base	N8010531		•	•	•					
96-Well Reaction Tube/Tray/Retainer Set, 0.2 mL	403083, 403086			•	•					

\* Multiple colors are available.

**Note:** Experiments using one or two 8-tube strips with attached caps require blank tube strips to balance lid pressure on the block or the use of the Applied Biosystems™ MicroAmp™ 96-Well Tray/Retainer Set (Cat. No. 4381850)—bottom part of the tray *only*; For use with the 96-well block of the ProFlex, SimpliAmp, Veriti, VeritiPro, MiniAmp Plus, and MiniAmp thermal cyclers.

## MicroAmp™ plastic consumables compatibility chart for Applied Biosystems™ real-time PCR systems

Product	Cat. No.	48-well		96-well 0.2 mL			96-well 0.1 mL			384-well
		StepOne™		7000	7300, 7500	QuantStudio™ 3, 5, 6, 6 Pro, 7, 7 Pro, 12K; ViiA™ 7; 7900HT	StepOnePlus™	7500	QuantStudio 3, 5, 6, 6 Pro, 7, 7 Pro, 12K; ViiA 7; 7900HT**	QuantStudio 5, 6, 6 Pro, 7, 7 Pro, 12K; ViiA 7; 7900HT
<b>96-well 0.2 mL reaction plates</b>										
Optical 96-Well Plate	N8010560, 4316813			•	•	•				
Optical 96-Well Plate with Barcode	4306737, 4326659			•	•	•				
Optical 96-Well Plate with Barcode & Optical Caps	403012			•	•	•				
Optical 96-Well Plate with Barcode & Optical Adhesive Films	4314320			•	•	•				
EnduraPlate Optical 96-Well Clear Plate	A36924				•	•				
EnduraPlate Optical 96-Well Clear Plate with Barcode*	4483354, 4483352				•	•				
<b>96-well 0.1 mL reaction plates</b>										
Fast Optical 96-Well Plate, 0.1 mL	4346907						•	•	•	
Fast Optical 96-Well Plate with Barcode, 0.1 mL	4346906, 4366932						•	•	•	
EnduraPlate Optical 96-Well Fast Plate	A36930						•	•	•	
EnduraPlate Optical 96-Well Fast Clear Plate with Barcode*	4483485, 4483494						•	•	•	
<b>384-well reaction plates</b>										
Optical 384-Well Plate	4343370									•
Optical 384-Well Plate with Barcode	4309849, 4326270, 4343814									•
EnduraPlate Optical 384-Well Plate	A36931									•
EnduraPlate Optical 384-Well Clear Plate with Barcode*	4483285, 4483273									•
<b>48-well reaction plates</b>										
Fast Optical 48-Well Plate	4375816	•								
<b>Strip tubes and caps</b>										
Fast 8-Tube Strip, 0.1 mL	4358293	•					•	•	•	
Optical 8-Tube Strip with Attached Optical Caps, 0.2 mL	A30588			•	•	•				
Optical 8-Tube Strip, 0.2 mL	4316567			•	•	•				
Optical 8-Cap Strip	4323032	•		•	•	•	•	•	•	
<b>Single tubes and caps</b>										
Fast Reaction Tube with Cap, 0.1 mL	4358297	•					•		•	
Optical Tube without Cap, 0.2 mL	N8010933			•	•					
<b>Seals and covers</b>										
Optical Adhesive Film	4360954, 4311971			•	•	•	•	•	•	•
48-Well Optical Adhesive Film	4375323	•								
<b>Reaction trays</b>										
96-Well Tray/Retainer Set	403081			•						
Fast 48-Well Tray	4375282	•								
96-Well Tray for VeriFlex Blocks	4379983						•			
<b>Accessories</b>										
Splash-Free 96-Well Base	4312063			•	•	•	•	•	•	
96-Well Support Base	4379590			•	•	•	•	•	•	
96-Well Base	N8010531			•	•	•	•	•	•	

\* Multiple colors are available.

\*\* 7900HT (0.2 mL and 0.1 mL) instruments are not recommended for use with individual tubes or tube strips.

**Note:** Experiments using one or two 8-tube strips with attached caps require blank tube strips to balance lid pressure on the block or the use of the MicroAmp 96-Well Tray/Retainer Set (Cat. No. 4381850)—bottom part of the tray *only*. For use with the 96-well block of the 7000, 7300, 7500, and ViiA 7 systems, and the QuantStudio 3, 5, 6, 6 Pro, 7, 7 Pro, and 12K instruments.

Find out more at [thermofisher.com/pcrplastics](https://thermofisher.com/pcrplastics)